



Dynalab Corp. Dynalon Labware

350 Commerce Drive
Rochester, New York 14623 USA

Tel: 585-334-2060 800-828-6595
Fax:585-334-0241
dynalabcorp.com dynalon.com

MANUFACTURED IN THE UK

TECHNICAL DATA SHEET

PRODUCT

CENTRIFUGE TUBES

SPECIFICATION

Sterility
Material tube
Material cap

Gamma Irradiated
Clear, non-toxic, medical grade polypropylene
Blue polyethylene

Manufacture

- 1 All polypropelene centrifuge tubes are manufactured in compliance with product and quality requirements as documented in the BS EN ISO 9001:2000 Quality System.
- 2 Verification of product clean room conditions in accordance with BS EN ISO 14644 'Environmental Cleanliness in enclosed spaces'
- 3 Sterilised by gamma irradiation (dosage 10 – 30 kGy)
- 4 Sterility is obtained through irradiation according to ISO 11137 (sterilisation of healthcare products - requirements for validation and routine control - radiation sterilisation) and EN552 (sterilisation of Medical Devices – validation and routine control of sterilisation by irradiation).

Item Code	Description	Capacity (ml)	H x OD (mm)	Bulk/Racked	Sterility	Material (Tube/Cap)	Inner Pack Qty	Case Qty
9936-0004	Centrifuge tube	15	120 x 17	Bulk	IRR	PP/PE	50	800
9936-0003	Centrifuge tube	15	120 x 17	Racked	IRR	PP/PE	50	450
15PSB	Centrifuge tube	15	120 x 17	Bulk	IRR	PS/PE	50	800
15PSR	Centrifuge tube	15	120 x 17	Racked	IRR	PS/PE	50	450
9936-0001	Centrifuge tube, skirted	50	115 x 31	-	IRR	PP/PE	25	250
9936-0002	Centrifuge tube, non-skirted	50	115 x 31	-	IRR	PP/PE	25	250



IWAKI

